



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/016,310	11/02/2001	Eric Brown	Brown.E-11	7835

22197 7590 02/06/2006

GENE SCOTT; PATENT LAW & VENTURE GROUP
3140 RED HILL AVENUE
SUITE 150
COSTA MESA, CA 92626-3440

EXAMINER

KOHARSKI, CHRISTOPHER

ART UNIT PAPER NUMBER

3763

DATE MAILED: 02/06/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/016,310

Applicant(s)

BROWN, ERIC

Examiner

Christopher D. Koharski

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 20 January 2006.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-16 is/are pending in the application.
- 4a) Of the above claim(s) 1-9 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 10-16 is/are rejected.
- 7) ☒ Claim(s) 10 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 02 November 2001 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date 11/02/2001.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____.

Art Unit: 3763

DETAILED ACTION

Election/Restrictions

Applicant's election without traverse of Group II (claims 10-16) in the reply filed on 1/05/2006 is acknowledged.

Claims 1-9 are withdrawn from further consideration pursuant to 37 CFR 1.142(b) as being drawn to a nonelected invention, there being no allowable generic or linking claim.

Information Disclosure Statement

The information disclosure statement (IDS) that was submitted on 11/02/2001 is in compliance with the provisions of 37 CFR 1.97. Accordingly, the examiner is considering the information disclosure statement.

Although it appears the reference Melcher (5,766,791) does not correlate with the scope of this invention, Applicant is requested to check this reference's relevance to this application.

Claim Objections

Claim 10 is objected to because of the following informalities: The reference to the "counter" in c) is a lack of antecedent basis issue, the term "counter" is not referenced previously. Examiner will interpret the "counter" to be related to the medical device means. Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this

Office action:

A person shall be entitled to a patent unless –

(a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.

Claims 10-14, and 16 rejected under 35 U.S.C 102(b) as being anticipated by Tacklind et al. (5,704,366). Tacklind et al. discloses a system for monitoring and reporting medical measurements.

Regarding claim 10, Tacklind et al. discloses a portable medical device with a sensing means (col 4, ln 15) with a recording and communication means (Figure 1). The medical state is reported by the medical sensing means (col 5, ln 21). The time and date are placed with the data sets (col 5 ln 15-38) and is capable of being configured with a wave energy means (col 11, ln 26).

Regarding claim 11, Tacklind et al. discloses a human response system (col 4, ln 25).

Regarding claims 12-13, Tacklind et al. discloses the medical information is transmitted from the unit to the base (col 7, ln 1; Figure 1).

Regarding claims 14 and 16, Tacklind et al. discloses a remote monitoring station that integrates medical data sets and interprets them (col 4, ln 10-30) and uses an error checking protocol (col 7, ln 27).

Art Unit: 3763

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claim 15 is rejected under 35 U.S.C. 103(a) as being unpatentable over Tacklind et al. in view of Ridgeway (5,967,975). Tacklind et al. meets the claim limitations as described above but does not include a timely transmission time between the base and unit.

However, Ridgeway teaches a home health parameter monitoring system. The reference teaches the use of transmission data sets that are sent under certain time periods to account for signal loss or equipment failure (col 8, ln 40-70).

At the time of the invention, it would have been obvious to use the repetitive transmission system of Ridgeway with the system of Tacklind et al. The addition of the data transmission provides the system with the ability to re-send data and maintain a link and keep trying in case the link is not established in a timely manner to maintain communications. Both references are analogous in the art and with the instant invention; therefore, a combination is proper. Therefore, one skilled in the art would have combined the teachings in the references in light of the disclosure of Ridgeway.

Art Unit: 3763

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Koharski whose telephone number is 571-272-7230. The examiner can normally be reached on Monday through Friday 7:30am-4:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

[Date]

1/24/06


Christopher Koharski
Examiner
Art Unit 3763